



LABORATORIO CHIMICO CAMERA DI COMMERCIO TORINO
Azienda Speciale della Camera di Commercio, Industria, Artigianato e Agricoltura di Torino

TURIN, 05/11/2012 DATE OF RECEIPT, 11/10/2012 C 5936/2012 Pagina 1 di 2

TO SGM LEKTRA SRL
VIA PAPA GIOVANNI XXIII, 49
RODANO MILANESE
64100 MI

TEST REPORT

It is forbidden to reproduce this test report partially.
Results are representative only of the sample tested.

SAMPLE: STEEL SENSOR FOR KPL LEVEL TRANSMITTER INSTRUMENT
WITH HYDROSTATIC STEAL

SAMPLER: CUSTOMER

CHEMICAL PARAMETERS

OVERALL MIGRATION IN DISTILLED WATER FOR 24 HOURS AT 40°C 2,3 mg/dm²

[METHOD: DM n° 174 6/04/2004 GU n° 166 17/07/2004]
Reference value : <=50 mg/dm²
Surface/volume ratio = 0,6
{Limits source: DM 06/04/2004 N° 174}
Start date test 29/10/2012 End date test: 30/10/2012

* SAMPLE PREPARATION

[METHOD: DM n° 338 22/07/1998 GU n° 228 30/09/1998 All pag 15 +
DM n° 220 26/04/1993 SO GU n° 162 13/07/1993 All III +
DM n° 123 28/03/2003 SO GU n° 125 31/05/2003 +
DM n° 299 22/12/2005 GU n° 37 14/02/2006 +
DM n° 174 24/09/2008 SO GU n° 261 07/11/2008]
Test date: 23/10/2012

* CHROME INTO THE MIGRATION LIQUID

N.D. less than 0,01 mg/kg

Test laboratory: Laboratorio Chimico Camera Commercio di Savona
[METHOD: UNI 12457-2:2004 + APAT 29/2003 - Met3150(2003)]
Reference value : <= 0,1 mg/kg
Start date test: 30/10/2012 End date test: 02/11/2012
{Limits source: DM 21/03/1973 (AGG. DM 258/2010)}

* NICKEL INTO THE MIGRATION LIQUID

N.D. less than 0,04 mg/kg

Test laboratory: Laboratorio Chimico Camera Commercio di Savona
[METHOD: UNI 12457-2:2004 + APAT 29/2003 - Met.3220(2003)]



LAB N° 0064

Via Ventimiglia, 165 - 10127 Torino
P. IVA 09273250010
Tel. 011.6700.111 - Fax. 011.6700.100
Email: labchim@lab-to.camcom.it
Sito Internet: www.lab-to.camcom.it



LABORATORIO CHIMICO CAMERA DI COMMERCIO TORINO
Azienda Speciale della Camera di Commercio, Industria, Artigianato e Agricoltura di Torino

TURIN, 05/11/2012

DATE OF RECEIPT, 11/10/2012

C 5936/2012

Pagina 2 di 2

TO

K.R.C.A. srl
Via S. Antonio Di Ranverso, 6
10090 ROSTA TO

[METHOD: UNI 12457-2:2004 + APAT 29/2003 - Met3220(2003)]

Reference value : $\leq 0,1$ mg/kg

Start date test: 30/10/2012 End date test: 02/11/2012

{Limits source: DM 21/03/1973 (AGG. DM 258/2010)}

* MANGANESE INTO THE MIGRATION LIQUID

0,054 mg/kg

Test laboratory: Laboratorio Chimico Camera Commercio di Savona

[METHOD: UNI 12457-2:2004 + APAT 29/2003 - Met3150(2003)]

Reference value : $\leq 0,1$ mg/kg


Start date test: 30/10/2012 End date test: 02/11/2012

{Limits source: DM 21/03/1973 (AGG. DM 258/2010)}

* ACCREDIA not accredited

In place of **RESPONSIBLE CHEMIST**

Palella dr.ssa Carola
Fracchia dr. Franco



LAB N° 0064

Via Ventimiglia, 165 - 10127 Torino
P. IVA 09273250010
Tel. 011.6700.111 - Fax. 011.6700.100
Email: labchim@lab-to.camcom.it
Sito Internet: www.lab-to.camcom.it